

導引用標準法碼最大允許誤差表

單位:mg

待檢定衡器/ 標稱值 自備法碼等級	Ⓐ /F <sub>1</sub>	Ⓑ /M <sub>1</sub>	Ⓒ /M <sub>2</sub>
1000 kg	5000	50000	160000
500 kg	2500	25000	80000
200 kg	1000	10000	30000
100 kg	500	5000	16000
50 kg	250	2500	8000
20 kg	100	1000	3000
10 kg	50	500	1600
5 kg	25	250	800
2 kg	10	100	300
1 kg	5	50	160
500 g	2.5	25	80
200 g	1.0	10	30
100 g	0.5	5.0	16
50 g	0.3	3.0	10
20 g	0.25	2.5	8.0
10 g	0.20	2.0	6.0
5 g	0.16	1.6	5.0
2 g	0.12	1.2	4.0
1 g	0.10	1.0	3.0

註：表中允許誤差係指正負誤差

檢定用法碼最大允許誤差表

單位:mg

待檢定衡器/ 標稱值 自備法碼等級	II /F <sub>2</sub>	III /M <sub>2</sub>	III /M <sub>3</sub>
1000 kg	16000	160000	500000
500 kg	8000	80000	250000
200 kg	3000	30000	100000
100 kg	1600	16000	50000
50 kg	800	8000	25000
20 kg	300	3000	10000
10 kg	160	1600	5000
5 kg	80	800	2500
2 kg	30	300	1000
1 kg	16	160	500
500 g	8.0	80	250
200 g	3.0	30	100
100 g	1.6	16	50
50 g	1.0	10	30
20 g	0.8	8.0	25
10 g	0.6	6.0	20
5 g	0.5	5.0	16
2 g	0.4	4.0	12
1 g	0.3	3.0	10

註：表中允許誤差係指正負誤差